



CLEARING MEMORY 920M+ SERIES OXIMETERS

1. Press and hold the small button while turning on the oximeter then release both buttons
2. Oximeter will read “CLr no”
3. Press small button
4. Display reads “CLr YES”
5. Press large button
6. Display reads “dEL no”
7. Press small button
8. Display reads “dEL YES”
9. Press large button
10. Display reads “dnE CLr”
11. Display then reads “- -“
12. Press and hold large button until oximeter powers off